

Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: BiCarb Type: Photometric Version: 1
 Number: 2002 Assay availability: Enabled Date: 18.02.2020
 Run controls for onboard reagents by: Lot Time: 11:01:08
 Operator: ADMIN

Reaction definition Reagent / Sample Validity checks

Reaction mode: End up

Primary Secondary Read times
 Wavelength: 404 / 548 Main: 32 -- 33
 Last required read: 33
 Absorbance range: 0.000 -- 2.400 Color correction: 3 -- 4
 Sample blank type: None

Assay 13 of 54

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: BiCarb Type: Photometric Version: 1
 Number: *** Assay availability: Enabled Date: 20.02.2020
 Run controls for onboard reagents by: Lot Time: 17:14:07
 Operator: FSE

Reaction definition Reagent / Sample Validity checks

Reagent: BiCarb
 Diluent: Saline
 Diluent dispense mode: Type 0
 Reagent volume: 160
 Water volume:
 Dispense mode: Type 2

Dilution name	Sample	Diluted sample	Diluent	Water	Dilution factor	Default dilution
Standart	1.6				1:1.00	<input checked="" type="radio"/>
					1:2.02	<input type="radio"/>
						<input type="radio"/>

Assay of

Done Cancel

?

Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: BiCarb Assay number: *** Date: 20.02.2020
 Calibration method: Linear Time: 17:14
 Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator set: BC-std
 Replicates: 2 [Range 1 - 3]

Calibrator level: Blank: Water
 Cal 1: BC-std1
 Cal 2: None
 Cal 3: None
 Cal 4: None
 Cal 5: None
 Cal 6: None

Concentration: 0.0000

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: BiCarb Assay number: *** Date: 20.02.2020
 Calibration method: Linear Time: 17:14
 Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator: BC-std

Calibrator level	Sample	Diluted sample	Diluent	Water
Blank: Water	1.6			
Cal 1: BC-std1	1.6			
Cal 2: None				
Cal 3: None				
Cal 4: None				
Cal 5: None				
Cal 6: None				

Done Cancel

?

Бикарбонат

NTB аналог NAD+



Параметры для ввода в программу анализатора Architect c8000 / c4000

Единицы измерения - ммоль/л; знаков после запятой 1 (**Configure result units**).

Диапазон метода: 4-50 ммоль/л (**Configure assay parameters\ Results**).

Референсный интервал: 22-29 ммоль/л (**Configure assay parameters\ Results**).

При наличии ЛИС, необязательно.

Диапазон нормальных значений указан в соответствии с рекомендациями
производителя реагентов.

Калибровка линейная, калибратор Bicarbonate standard, нулевой - вода (или физраствор).

Контроль по TruLab Bicarbonate.