

Ревматоидный фактор. RF

ImmunoTD (Иммунотурбидиметрический)



Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

☒ General ☐ Calibration ☐ SmartWash ☐ Results ☐ Interpretation

Assay: RF Type: Photometric Version: 1
Number: *** Assay availability: Enabled Date: 18.02.2020
Run controls for onboard reagents by: Lot Time: 11:01:08
Operator: ADMIN

☒ Reaction definition ☐ Reagent / Sample ☐ Validity checks

Reaction mode: End up

Primary Secondary Read times
Wavelength: 340 / 804 Main: 32 -- 33
Last required read: 33
Absorbance range: -0.1 -- 2.500 Color correction: --
Sample blank type: Self Blank: 14 -- 16

Assay 13 of 54

Done Cancel

?

Configure assay parameters...

☒ General ☐ Calibration ☐ SmartWash ☐ Results ☐ Interpretation

Assay: RF Type: Photometric Version: 1
Number: *** Assay availability: Enabled Date: 20.02.2020
Run controls for onboard reagents by: Lot Time: 17:14:07
Operator: FSE

☐ Reaction definition ☒ Reagent / Sample ☐ Validity checks

Reagent: RF
Diluent: Saline
Diluent dispense mode: Type 0

R1 R2
Reagent volume: 150 30
Water volume: --
Dispense mode: Type 2 Type 2

Dilution name	Sample	Diluted sample	Diluent	Water	Dilution factor	Default dilution
Standart	9.0				= 1:1.00	<input checked="" type="radio"/>
1:3	3.0				= 1:3.00	<input type="radio"/>
1.5x	13.0				= 1.5:1.00	<input type="radio"/>

Assay of

Done Cancel

?

Ревматоидный фактор. RF

ImmunoTD (Иммунотурбидиметрический)



Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: RF Assay number: *** Date: 20.02.2020
Calibration method: Spline Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator set: TruCalRF
Calibrator level: Blank: Water Concentration: 0.0000
Cal 1: TruCalRF1
Cal 2: TruCalRF2
Cal 3: TruCalRF3
Cal 4: TruCalRF4
Cal 5: TruCalRF5
Cal 6: None

Replicates: 2 [Range 1 - 3]

Assay of

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: RF Assay number: *** Date: 20.02.2020
Calibration method: Spline Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator: TruCalRF

Calibrator level	Sample	Diluted sample	Diluent	Water
Blank: Water	9.0			
Cal 1: TruCalRF1	9.0			
Cal 2: TruCalRF2	9.0			
Cal 3: TruCalRF3	9.0			
Cal 4: TruCalRF4	9.0			
Cal 5: TruCalRF5	9.0			
Cal 6: None				

Assay of

Done Cancel

?

Ревматоидный фактор. RF

ImmunoTD (Иммунотурбидиметрический)



Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: RF Assay number: *** Date: 20.02.2020
Calibration method: Spline Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibration intervals:

Full interval: 0 (hours)
Adjust interval: 0 (hours)

Calibration type:

Adjust type: 2 Point
Adjust level: Cal 1
Default ordering type: Full

Assay of

Done
Cancel

?

Единицы измерения - МЕ/мл; знаков после запятой 2 (**Configure result units**).

Диапазон метода: 5-500 МЕ/мл (**Configure assay parameters\ Results**).

Рефереренсный интервал: 0-15 МЕ/мл (**Configure assay parameters\ Results**).

При наличии ЛИС, не обязательно.

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов.

Калибратор TruCal RF, набор из 5 уровней, нулевой - вода (или физраствор).

Контроль по TruLab protein уровень 1 и уровень 2.