

Параметры для ввода в программу анализатора Architect с8000 / с4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: Transf Type: Photometric Version: 1
 Number: *** Assay availability: Enabled Date: 18.02.2020
 Run controls for onboard reagents by: Lot Time: 11:01:08
 Operator: ADMIN

Reaction definition Reagent / Sample Validity checks

Reaction mode: End up

Primary Secondary Read times
 Wavelength: 572 / None Main: 32 -- 33
 Last required read: 33
 Absorbance range: 0.000 -- 2.500 Color correction: --
 Sample blank type: Self Blank: 14 -- 16

Assay 13 of 54

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: Transf Type: Photometric Version: 1
 Number: *** Assay availability: Enabled Date: 20.02.2020
 Run controls for onboard reagents by: Lot Time: 17:14:07
 Operator: FSE

Reaction definition Reagent / Sample Validity checks

Reagent: Transf
 Diluent: Saline
 Diluent dispense mode: Type 0

R1 R2
 Reagent volume: 190 38
 Water volume: --
 Dispense mode: Type 2 Type 2

Dilution name	Sample	Diluted sample	Diluent	Water	Dilution factor	Default dilution
Standart	1.5				1:1.00	<input checked="" type="radio"/>
1:4	20.0	1.5	80		1:4.00	<input type="radio"/>
						<input type="radio"/>

Assay of

Done Cancel

?

Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: Transf Assay number: *** Date: 20.02.2020
 Calibration method: Spline Time: 17:14
 Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator set: TruCalProt

Calibrator level: Blank: Water Concentration: 0.0000

Cal 1: TruCalProt1
 Cal 2: TruCalProt2
 Cal 3: TruCalProt3
 Cal 4: TruCalProt4
 Cal 5: TruCalProt5
 Cal 6: None

Replicates: 2 [Range 1 - 3]

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: Transf Assay number: *** Date: 20.02.2020
 Calibration method: Spline Time: 17:14
 Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator: TruCalProt

Calibrator level	Sample	Diluted sample	Diluent	Water
Blank: Water	1.5			
Cal 1: TruCalProt1	1.5			
Cal 2: TruCalProt2	1.5			
Cal 3: TruCalProt3	1.5			
Cal 4: TruCalProt4	1.5			
Cal 5: TruCalProt 5	1.5			
Cal 6: None				

Done Cancel

?

Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: Transf Assay number: *** Date: 20.02.2020
Calibration method: Spline Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibration intervals:

Full interval: 0 (hours)
Adjust interval: 0 (hours)

Calibration type:

Adjust type: 2 Point
Adjust level: Cal 1
Default ordering type: Full

Assay of

Done
Cancel

?

Единицы измерения - г/л; знаков после запятой 2 (Configure result units).

Диапазон метода: 0.03 - 8.0 г/л (Configure assay parameters\ Results).

Референсный интервал: 2.0-3.6 г/л (Configure assay parameters\ Results).

При наличии ЛИС, не обязательно.

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов.

Калибратор TruCal Protein, набор из 5 уровней, нулевой - вода (или физраствор).

Контроль по TruLab Protein уровень 1 и уровень 2.